

Dispositivos Médicos



BATA HOSPITALAR DESCARTÁVEL 20G

Rua Alexandre Brito Gouveia, s/n
Gavinhos de Baixo, 3400-004
Oliveira do Hospital, Portugal

Telf.: + 351 238 609 097

Tel.: +351 969 990 099

geral@wellgiven.com

www.wellgiven.com


BATA HOSPITALAR DESCARTÁVEL 20 G



Artigo	Nome		
GROW20G	Hospital Gown		
Descrição do Produto			
Composição: 100% polipropileno			
TNT 20 gr/m2			
Repelente e não impermeável			
Não esterilizada			
Manga comprida, fixa-se com fitas na parte de trás do pescoço e amarra-se na frente com comprimento para cruzar atrás e dar o nó na frente.			
Cores Disponíveis			
Azul	Amarelo	Verde	Branco

Cumpe todos os requisitos essenciais inerentes aos dispositivos médicos destinados a impedir a transmissão de agentes infecciosos entre profissionais de saúde e doentes durante procedimentos cirúrgicos e outros procedimentos, em ambientes hospitalares e afins.

MEDIDAS / EMBALAGEM



	MEDIDAS	CM
1	Altura	106
2	Comprimento da manga	55

INSTRUÇÕES DE LAVAGEM

Não aplicável. Descartável.

EMBALAGEM

10 Peças por embalagem.
110 Peças por caixa.
Tamanho da caixa 60x40x30



CE ATTESTATION OF CONFORMITY

Medical Device – Ref.: SURGICALGROWGazul
SURGICALGROWGamarelo
SURGICALGROWGverde
SURGICALGROWGbranco

Classification: Class I – Non Sterile

Device Description: Surgical Gown

Standars Applied: EN 13795-1:2019

EN 13795-2:2019

EN ISO 15797:2018

IEC 62366

ISO 14971

Manufactured by:

Wellgiven, Lda

Rua Alexandre Brito Gouveia, s/n

Gavinhos de Baixo, 3400-004

Oliveira do Hospital, Portugal

We, MANUFACTUER, declare under our sole responsibility, that Medical Device is in conformity with the requirements of the Annex VII of MDD 93/42/CEE council directive including the other applicable directives/regulations.

Sandra Santos

The devices it manufactures, Disposable Surgical Gown (Refs: GROW20Gazul, GROW20Gamarelo; GROW20Gverde; GROW20Gbranco), comply with the essential requirements established in Annex I of Directive 93/42 / EEC, of 14 June, in its current wording and Decree- Law No. 145/2009, of 17 June, which is applicable to them, so it does not compromise the clinical status or patient safety, nor, yet, the safety and health of users or, possibly, third parties, when used under the conditions and for the intended purposes, considering that the possible risks associated with the intended use are acceptable risks when compared to the benefit provided to patients and are compatible with a high degree of health and safety protection.