

Dispositivos Médicos



BATA HOSPITALAR DESCARTÁVEL 30G

Rua Alexandre Brito Gouveia, s/n
Gavinhos de Baixo, 3400-004
Oliveira do Hospital, Portugal

Tel.: + 351 238 609 097

Tel.: +351 969 990 099

geral@wellgiven.com

www.wellgiven.com


BATA HOSPITALAR DESCARTÁVEL 30G



| Artigo | Nome | | |
|---|---------------|-------|--------|
| GROW30GLAM | Hospital Gown | | |
| Descrição do Produto | | | |
| Composição: 50% TNT 50% Laminada | | | |
| TNT Laminado 30 gr/m ² | | | |
| Impermeável Laminada | | | |
| Não esterilizada | | | |
| Manga comprida, fixa-se com fitas na parte de trás do pescoço e amarra-se na frente com comprimento para cruzar atrás e dar o nó na frente. | | | |
| Cores Disponíveis | | | |
| Azul | Amarelo | Verde | Branco |

Cumpra todos os requisitos essenciais inerentes aos dispositivos médicos destinados a impedir a transmissão de agentes infecciosos entre profissionais de saúde e doentes durante procedimentos cirúrgicos e outros procedimentos, em ambientes hospitalares e afins.

MEDIDAS / EMBALAGEM



| MEDIDAS | | CM |
|---------|----------------------|-----|
| 1 | Altura | 106 |
| 2 | Comprimento da manga | 56 |

INSTRUÇÕES DE LAVAGEM
Não aplicável. Descartável.

EMBALAGEM
10 Peças por embalagem.
100 Peças por caixa.
Tamanho da caixa 60x40x40



CE ATTESTATION OF CONFORMITY

Medical Device – Ref.: GROW30GLAMblue
GROW30GLAMyellow
GROW30GLAMgreen
GROW30GLAMwhite

Classification: Class I – Non Sterile


Device Description: Hospital Gown

Standars Applied: EN 13795-1:2019
EN 13795-2:2019
EN ISO 15797:2018
IEC 62366
ISO 14971

Manufactured by:

Wellgiven, Lda
Rua Alexandre Brito Gouveia, s/n
Gavinhos de Baixo, 3400-004
Oliveira do Hospital, Portugal

We, MANUFACTUER, declare under our sole responsibility, that Medical Device is in conformity with the requirements of the Annex VII of MDD 93/42/CEE council directive including the other applicable directives/regulations.



The devices it manufactures, Disposable Surgical Gown (Refs: GROW30GLAMblue, GROW30GLAMyellow; GROW30GLAMgreen; GROW30GLAMwhite), comply with the essential requirements established in Annex I of Directive 93/42 / EEC, of 14 June, in its current wording and Decree- Law No. 145/2009, of 17 June, which is applicable to them, so it does not compromise the clinical status or patient safety, nor, yet, the safety and health of users or, possibly, third parties, when used under the conditions and for the intended purposes, considering that the possible risks associated with the intended use are acceptable risks when compared to the benefit provided to patients and are compatible with a high degree of health and safety protection.